

Certificate of Compliance

Certificate: 2300657 (LR 96902)

Master Contract: 172608

Project: 2435658

Date Issued: July 4, 2011

Issued to: Yokogawa Electric Corporation
2-9-32 Nakacho
Musashino-shi, Tokyo 180-8750
Japan
Attention: Junji Takano

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Constantine Somow

Issued by: Constantine Somow, CTech

PRODUCTS

- CLASS 3631 05** - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT
CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Two-wire analyzer, Model FLXA21* Series, K9698* Series, rated 24 V dc, 4-20 mA, Type 3S/4X.

Note: The above model is Pollution Degree 2, Installation Category I, powered from SELV supply.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (2nd Edition) - Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Second Edition)



Supplement to Certificate of Compliance

Certificate: 2300657

Master Contract: 172608

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
2435658	July 4, 2011	Update Report 2300657 to include an alternative construction.
2300657	October 20, 2010	Two-wire analyzer, Model FLXA21* Series, K9698* Series (C/US).